

Application Note

Document Number APL-00152

Updated 05/03/02 Rev-A



New NAAI SCPI Command

Symptom

If you have connected to the NAAI via the serial port first, you may not be able to connect using Ethernet without power-cycling the unit.

A new SCPI command was developed so that the power cycling would not be required. Issue the following SCPI command on the serial port

Syst:rethernet

This command sets the remote port to be the Ethernet port and therefor allows remote control of the test set over the Ethernet port.